

TEST REPORT	
EN 420	
Protective gloves - General requirements and test methods	
Report Number.....	BTa2003240120SR
Tested by (+ signature)	Alice Luo
Compiled by (+ signature)	Jane Lin
Approved by (+ signature)	Jacky Wu
Date of issue.....	Apr. 08, 2020
Total number of pages.....	17 pages
Testing Laboratory	Shenzhen BTa Product Testing Co., Ltd.
Address.....	Room410, Floor 4, new tea village industrial park gushu, community Xixiang streetbaoan district. Shenzhen China
Address.....	As above
Applicant's name	Relabeled for Innovation Line
Test specification:	
Standard	EN 420:2003+A1:2009
Test procedure.....	CE
Non-standard test method.....	N/A
Test Report Form No.	EN 420:2003+A1:2009
Test Report Form(s) Originator	BTa
Master TRF	Dated 2020-02
The test results presented in this report relate only to the object tested. This report shall not be reproduced except in full, without the written approval of the Laboratory. The authenticity of this Test Report and its contents can be verified by contacting the Laboratory, responsible for this Test Report.	
Test item description.....	NG100
Manufacturer.....	Relabeled for Innovation Line

This Test Report is issued by the Company subject to its General Conditions of Service printed overleaf. Attention is drawn to the limitations of liability, indemnification and jurisdictional policies defined therein. The results shown in this test report refer only to the sample(s) tested unless otherwise stated and the sample(s) are retained for 30 days only.

List of Attachments (including a total number of pages in each attachment):

-- Attachment 1: One pages for Photo documentation.

Summary of testing:**Tests performed (name of test and test clause):**

-- EN 420:2003+A1:2009

Medical face masks - Requirements and test methods

Testing location:Room410, Floor 4, new tea village industrial park
gushu, community Xixiang streetbaoan district.
Shenzhen China**Copy of marking plate**

The artwork below may be only a draft. The use of certification marks on a product must be authorized by the respective NCBs that own these marks.

(Additional requirements for markings. See 1.7 NOTE)

YXD

Disposable sanitary inspection gloves

YXD-001

EN 420:2003+A1:2009



**Possible test case verdicts:**

- test case does not apply to the test object.....: N/A(or N)
- test object does meet the requirement.....: P (Pass)
- test object does not meet the requirement.....: F (Fail)

Testing**Date of receipt of test item**: Feb. 23, 2020**Date (s) of performance of tests**: Feb. 23, 2020 – Mar.27, 2020**General remarks:****Throughout this report a ☒ comma / ☐ point is used as the decimal separator.****General product information:**



Requirements

Table 2 – Requirements for EN 420:2003+ A1:2009 Clause 5 Size and Dexterity

Glove size	6	7	8	9	10	11
Minimum length / mm	220	230	240	250	260	270

Performance level	1	2	3	4	5
Diameter of dexterity pin /mm	11	9.5	8	6.5	5

Test Results

Clause / Test	Requirement	Test Results			UoM (See note C)	Result
5.1 Glove length, comfort and fit	See table 2	Size	Length /mm		± 0.3 mm	PASS
			Left	Right		
		7	239	238		
		Comments on fit: Satisfactory				
		8	246	249		
		Comments on fit: Satisfactory				
		9	258	257		
		Comments on fit: Satisfactory				
		10	266	268		
		Comments on fit: Satisfactory				
		11	277	279		
		Comments on fit: Satisfactory				
5.2 Dexterity	See table 2	Size	Minimum pin diameter / mm		--	Level 5
		7	5.0			
		8	5.0			
		9	5.0			
		10	5.0			
		11	5.0			

Photo-documentation



===== End of test report =====